

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041522\
 Data File : VW025596.D
 Acq On : 16 Apr 2022 03:25
 Operator : SY/MD
 Sample : N2438-17
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6Z4

Quant Time: Apr 16 04:51:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 16 04:36:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	118003	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	120938	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	54396	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	39921	3.639	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.800%
7) Chloroethane-d5	1.571	69	43132	3.696	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.000%
11) 1,1-Dichloroethene-d2	2.111	63	66084	3.007	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.200%
20) 2-Butanone-d5	3.905	46	77069	57.751	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.500%
24) Chloroform-d	4.352	84	78323	4.724	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
26) 1,2-Dichloroethane-d4	5.037	65	37905	5.306	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.200%
32) Benzene-d6	5.053	84	141085	4.032	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.600%
36) 1,2-Dichloropropane-d6	6.072	67	45252	4.637	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.800%
41) Toluene-d8	7.316	98	115263	3.550	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.000%
43) trans-1,3-Dichloroprop...	7.628	79	10995	3.352	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.000%
46) 2-Hexanone-d5	8.088	63	56265	48.875	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.740%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	31342	5.084	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	42316	4.640	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.800%
Target Compounds						
					Qvalue	
13) Acetone	2.195	43	3740	4.129	ug/L	75
16) Methylene chloride	2.513	84	3892	0.392	ug/L	92
25) Chloroform	4.378	83	7655	0.499	ug/L	86
38) Bromodichloromethane	6.529	83	1050	0.102	ug/L #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041522\
Data File : VV025596.D
Acq On : 16 Apr 2022 03:25
Operator : SY/MD
Sample : N2438-17
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6Z4

Quant Time: Apr 16 04:51:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 16 04:36:16 2022
Response via : Initial Calibration

