

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW041524\  
 Data File : VW035209.D  
 Acq On : 15 Apr 2024 20:06  
 Operator : SY/MD  
 Sample : P2057-21  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VHBLK001

Quant Time: Apr 16 01:49:23 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR041524WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Tue Apr 16 00:33:27 2024  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.535  | 114   | 184619   | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.786  | 117   | 190386   | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.185 | 152   | 81004    | 5.000    | ug/L  | 0.00     |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.278  | 65    | 87614    | 5.068    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 101.400% |
| 7) Chloroethane-d5            | 1.532  | 69    | 71230    | 5.303    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 106.000% |
| 11) 1,1-Dichloroethene-d2     | 2.057  | 65    | 37923    | 5.171    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 103.400% |
| 20) 2-Butanone-d5             | 3.828  | 46    | 161647   | 55.284   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 110.560% |
| 24) Chloroform-d              | 4.249  | 84    | 139841   | 5.174    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 103.400% |
| 26) 1,2-Dichloroethane-d4     | 4.944  | 65    | 72095    | 5.604    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 112.000% |
| 32) Benzene-d6                | 4.960  | 84    | 275804   | 5.099    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 102.000% |
| 36) 1,2-Dichloropropane-d6    | 5.989  | 67    | 90325    | 5.370    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 107.400% |
| 41) Toluene-d8                | 7.243  | 98    | 227186   | 4.677    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 93.600%  |
| 43) trans-1,3-Dichloroprop... | 7.561  | 79    | 26612    | 4.433    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 88.600%  |
| 46) 2-Hexanone-d5             | 8.031  | 63    | 116774   | 46.704   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 93.400%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.153 | 84    | 64611    | 4.987    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 99.800%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.561 | 152   | 79369    | 5.512    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 110.200% |
| Target Compounds              |        |       |          |          |       |          |
| 16) Methylene chloride        | 2.449  | 84    | 5123     | 0.290    | ug/L  | 95       |
| -----                         |        |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV041524\  
Data File : VV035209.D  
Acq On : 15 Apr 2024 20:06  
Operator : SY/MD  
Sample : P2057-21  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VHBLK001

Quant Time: Apr 16 01:49:23 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR041524WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Tue Apr 16 00:33:27 2024  
Response via : Initial Calibration

