

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041524\
 Data File : VW035214.D
 Acq On : 15 Apr 2024 22:06
 Operator : SY/MD
 Sample : P2032-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EOMF4

Quant Time: Apr 16 01:50:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 16 00:33:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	182563	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	186344	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	81235	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	90010	5.265	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.400%
7) Chloroethane-d5	1.532	69	69473	5.231	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.600%
11) 1,1-Dichloroethene-d2	2.057	65	37998	5.240	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	104.800%
20) 2-Butanone-d5	3.809	46	186227	64.408	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	128.820%
24) Chloroform-d	4.249	84	142008	5.313	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.200%
26) 1,2-Dichloroethane-d4	4.944	65	72088	5.667	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.400%
32) Benzene-d6	4.960	84	276284	5.219	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.400%
36) 1,2-Dichloropropane-d6	5.989	67	90746	5.512	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.200%
41) Toluene-d8	7.243	98	225979	4.753	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.000%
43) trans-1,3-Dichloroprop...	7.561	79	27206	4.630	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.600%
46) 2-Hexanone-d5	8.031	63	121519	49.656	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.320%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	68285	5.385	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.600%
66) 1,2-Dichlorobenzene-d4	11.561	152	79573	5.511	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.200%
Target Compounds						
5) Vinyl chloride	1.282	62	61971	3.134	ug/L	98
12) 1,1-Dichloroethene	2.066	96	4545	0.328	ug/L	# 1
18) trans-1,2-Dichloroethene	2.696	96	40196	2.838	ug/L	97
22) cis-1,2-Dichloroethene	3.809	96	1608968	110.501	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041524\
Data File : VV035214.D
Acq On : 15 Apr 2024 22:06
Operator : SY/MD
Sample : P2032-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EOMF4

Quant Time: Apr 16 01:50:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR041524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 16 00:33:27 2024
Response via : Initial Calibration

