

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041723\
 Data File : VW030586.D
 Acq On : 17 Apr 2023 14:12
 Operator : SY/MD
 Sample : VW0417WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK321

Quant Time: Apr 18 02:48:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 18 02:45:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	262775	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	254990	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	123746	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	85520	44.395	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.780%
7) Chloroethane-d5	1.539	69	69535	43.075	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.160%
11) 1,1-Dichloroethene-d2	2.066	65	40426	44.252	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.500%
21) 2-Butanone-d5	3.806	46	101775	96.190	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.190%
24) Chloroform-d	4.262	84	148566	42.557	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.120%
26) 1,2-Dichloroethane-d4	4.953	65	103058	45.917	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.840%
32) Benzene-d6	4.970	84	281385	46.674	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.340%
36) 1,2-Dichloropropane-d6	5.998	67	88356	47.182	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.360%
41) Toluene-d8	7.252	98	251738	44.474	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.940%
43) trans-1,3-Dichloroprop...	7.561	79	44324	46.960	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.920%
47) 2-Hexanone-d5	8.034	63	64370	87.256	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.260%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	117486	43.236	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.480%
66) 1,2-Dichlorobenzene-d4	11.571	152	98743	50.202	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041723\
Data File : VV030586.D
Acq On : 17 Apr 2023 14:12
Operator : SY/MD
Sample : VV0417WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK321

Quant Time: Apr 18 02:48:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 18 02:45:56 2023
Response via : Initial Calibration

