

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041723\
 Data File : VW030593.D
 Acq On : 17 Apr 2023 16:59
 Operator : SY/MD
 Sample : 02360-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COAN3

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/18/2023
 Supervised By : Mahesh Dadoda 04/18/2023

Quant Time: Apr 18 02:50:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 18 02:45:56 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	215467	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	206212	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	101240	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	80123	50.725	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 101.460%			
7) Chloroethane-d5	1.535	69	65613	49.570	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 99.140%			
11) 1,1-Dichloroethene-d2	2.066	65	37544	50.121	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 100.240%			
21) 2-Butanone-d5	3.802	46	92208	106.282	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 106.280%			
24) Chloroform-d	4.259	84	142714	49.856	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 99.720%			
26) 1,2-Dichloroethane-d4	4.953	65	97213	52.823	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 105.640%			
32) Benzene-d6	4.973	84	267296	54.825	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 109.660%			
36) 1,2-Dichloropropane-d6	6.008	67	84603	55.864	ug/L	0.01
Spiked Amount 50.000	Range 70 - 120		Recovery = 111.720%			
41) Toluene-d8	7.252	98	230725	50.403	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 100.800%			
43) trans-1,3-Dichloroprop...	7.561	79	38453	50.377	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 100.760%			
47) 2-Hexanone-d5	8.034	63	58045	97.294	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 97.290%			
56) 1,1,2,2-Tetrachloroeth...	10.162	84	104551	47.577	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 95.160%			
66) 1,2-Dichlorobenzene-d4	11.570	152	87774	54.546	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 109.100%			
Target Compounds						Qvalue
10) 1,1,2-Trichloro-1,2,2-...	2.076	101	53540	32.084	ug/L	99
12) 1,1-Dichloroethene	2.076	96	25885	17.157	ug/L #	31
17) trans-1,2-Dichloroethene	2.706	96	1434	0.942	ug/L	92
19) 1,1-Dichloroethane	3.124	63	4485	1.444	ug/L #	88
20) cis-1,2-Dichloroethene	3.825	96	121153	73.123	ug/L	98
25) Chloroform	4.288	83	11551	3.594	ug/L	93
30) 1,1,1-Trichloroethane	4.522	97	159751	55.132	ug/L	99
31) Carbon tetrachloride	4.751	117	8305	3.206	ug/L	97
33) Benzene	5.024	78	6704	1.071	ug/L	100
34) Trichloroethene	5.847	95	35342070m	21698.069	ug/L	
45) 1,1,2-Trichloroethane	7.780	97	12301	7.278	ug/L	95
46) Tetrachloroethene	7.915	164	6693	4.694	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041723\
Data File : VV030593.D
Acq On : 17 Apr 2023 16:59
Operator : SY/MD
Sample : 02360-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AN3

Manual Integrations
APPROVED

Reviewed By : John Carlone 04/18/2023
Supervised By : Mahesh Dadoda 04/18/2023

Quant Time: Apr 18 02:50:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 18 02:45:56 2023
Response via : Initial Calibration

