

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041824\
 Data File : VV035272.D
 Acq On : 18 Apr 2024 10:42
 Operator : SY/MD
 Sample : VV0418WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK269

Quant Time: Apr 19 01:41:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 10 04:53:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	354304	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	344328	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	149918	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	166655	48.705	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.420%
7) Chloroethane-d5	1.523	69	129226	43.481	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.960%
11) 1,1-Dichloroethene-d2	2.056	65	72659	48.442	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.880%
21) 2-Butanone-d5	3.786	46	225075	110.910	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.910%
24) Chloroform-d	4.246	84	312197	54.267	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.540%
26) 1,2-Dichloroethane-d4	4.937	65	203124	55.612	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.220%
32) Benzene-d6	4.957	84	628917	54.376	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.760%
36) 1,2-Dichloropropane-d6	5.986	67	207632	56.532	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	113.060%
41) Toluene-d8	7.240	98	535835	51.442	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.880%
43) trans-1,3-Dichloroprop...	7.551	79	84608	47.386	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.780%
47) 2-Hexanone-d5	8.024	63	140580	101.236	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.240%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	274790	54.331	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	108.660%
66) 1,2-Dichlorobenzene-d4	11.558	152	191606	56.924	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041824\
Data File : VV035272.D
Acq On : 18 Apr 2024 10:42
Operator : SY/MD
Sample : VV0418WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK269

Quant Time: Apr 19 01:41:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 10 04:53:36 2024
Response via : Initial Calibration

