

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041824\
 Data File : VV035292.D
 Acq On : 18 Apr 2024 18:50
 Operator : SY/MD
 Sample : P2157-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CORJ7

Integration Parameters: LSCINT.P

Integrator: RTE

Smoothing : OFF

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 0 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Title : VOC Analysis

Signal : TIC: VV035292.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.278	68	74	89	rVB	202543	219946	13.70%	1.842%
2	1.523	144	150	168	rVB	182050	231043	14.39%	1.935%
3	2.056	307	316	330	rBV	350989	500944	31.20%	4.196%
4	2.899	575	578	580	rVB2	274	119	0.01%	0.001%
5	3.034	616	620	623	rBV3	261	239	0.01%	0.002%
6	3.166	659	661	664	rVB2	440	229	0.01%	0.002%
7	3.253	686	688	690	rBV3	199	129	0.01%	0.001%
8	3.362	720	722	724	rBV2	210	78	0.00%	0.001%
9	3.555	781	782	784	rBV	182	97	0.01%	0.001%
10	3.603	793	797	800	rBV2	335	201	0.01%	0.002%
11	3.619	800	802	805	rBV3	160	71	0.00%	0.001%
12	3.635	805	807	808	rBV	167	50	0.00%	0.000%
13	3.667	814	817	819	rBV2	311	215	0.01%	0.002%
14	3.703	825	828	829	rBV	297	180	0.01%	0.002%
15	3.754	840	844	846	rBV2	297	178	0.01%	0.001%
16	3.786	846	854	906	rBV	117869	352911	21.98%	2.956%
17	4.246	983	997	1028	rBV	249563	640323	39.88%	5.363%
18	4.738	1144	1150	1158	rBV7	1744	2614	0.16%	0.022%
19	4.786	1163	1165	1169	rBV3	418	262	0.02%	0.002%
20	4.953	1201	1217	1251	rBV2	602886	1605446	100.00%	13.447%
21	5.281	1317	1319	1323	rBV2	527	338	0.02%	0.003%
22	5.378	1347	1349	1353	rVB2	513	255	0.02%	0.002%
23	5.497	1382	1386	1387	rBV2	643	425	0.03%	0.004%
24	5.532	1387	1397	1433	rVV	357146	753184	46.91%	6.308%
25	5.834	1482	1491	1493	rBV5	4965	6192	0.39%	0.052%
26	5.986	1526	1538	1571	rBV	350609	733419	45.68%	6.143%
27	6.404	1664	1668	1671	rBV2	371	219	0.01%	0.002%
28	6.465	1685	1687	1695	rVB2	664	430	0.03%	0.004%
29	6.516	1699	1703	1706	rBV2	359	200	0.01%	0.002%
30	6.561	1714	1717	1718	rBV	186	82	0.01%	0.001%
31	6.593	1724	1727	1728	rBV	174	88	0.01%	0.001%
32	6.603	1728	1730	1731	rBV	304	104	0.01%	0.001%
33	6.699	1758	1760	1762	rVB2	219	82	0.01%	0.001%
34	6.738	1768	1772	1774	rBV3	453	323	0.02%	0.003%
35	6.870	1809	1813	1816	rBV3	436	443	0.03%	0.004%
36	6.912	1816	1826	1854	rBV	167769	327505	20.40%	2.743%

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041824\
 Data File : VV035292.D
 Acq On : 18 Apr 2024 18:50
 Operator : SY/MD
 Sample : P2157-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CORJ7

Integration Parameters: LSCINT.P

Integrator: RTE

Smoothing : OFF

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 0 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Title : VOC Analysis

37	7.243	1918	1929	1967	rBV	659209	1190230	74.14%	9.969%
38	7.555	2015	2026	2052	rBV	116379	221064	13.77%	1.852%
39	7.857	2118	2120	2124	rBV3	614	432	0.03%	0.004%
40	7.915	2135	2138	2147	rVB5	1609	1769	0.11%	0.015%
41	7.979	2156	2158	2161	rBV2	460	328	0.02%	0.003%
42	8.024	2163	2172	2215	rBV	349041	728259	45.36%	6.100%
43	8.783	2396	2408	2443	rBV2	555288	1128772	70.31%	9.454%
44	9.461	2617	2619	2622	rBV	266	130	0.01%	0.001%
45	9.529	2638	2640	2642	rBV2	232	115	0.01%	0.001%
46	9.590	2655	2659	2662	rBV4	442	321	0.02%	0.003%
47	9.606	2662	2664	2666	rBV2	235	96	0.01%	0.001%
48	9.638	2672	2674	2676	rBV2	298	100	0.01%	0.001%
49	9.760	2706	2712	2715	rBV2	434	410	0.03%	0.003%
50	9.783	2715	2719	2720	rBV2	216	117	0.01%	0.001%
51	9.866	2742	2745	2747	rBV3	807	595	0.04%	0.005%
52	9.937	2764	2767	2769	rBV	314	198	0.01%	0.002%
53	10.014	2787	2791	2794	rBV2	345	182	0.01%	0.002%
54	10.101	2816	2818	2822	rBV	192	164	0.01%	0.001%
55	10.149	2822	2833	2860	rBV	499768	800655	49.87%	6.706%
56	10.368	2899	2901	2904	rBV4	296	154	0.01%	0.001%
57	10.413	2913	2915	2919	rBV2	299	261	0.02%	0.002%
58	10.493	2937	2940	2942	rBV2	439	223	0.01%	0.002%
59	10.535	2949	2953	2957	rBV2	314	340	0.02%	0.003%
60	10.583	2965	2968	2970	rBV2	308	197	0.01%	0.002%
61	10.635	2982	2984	2990	rVB	370	226	0.01%	0.002%
62	10.670	2990	2995	3003	rBV4	583	799	0.05%	0.007%
63	10.702	3003	3005	3006	rBV4	243	94	0.01%	0.001%
64	10.735	3012	3015	3018	rVB	344	192	0.01%	0.002%
65	10.747	3018	3019	3020	rBV	152	42	0.00%	0.000%
66	10.815	3037	3040	3042	rBV	289	175	0.01%	0.001%
67	10.831	3042	3045	3046	rBV	253	162	0.01%	0.001%
68	10.873	3051	3058	3064	rBV4	842	1243	0.08%	0.010%
69	11.046	3111	3112	3114	rBV	324	123	0.01%	0.001%
70	11.098	3125	3128	3129	rBV	364	139	0.01%	0.001%
71	11.127	3129	3137	3144	rBV3	10339	16301	1.02%	0.137%
72	11.185	3145	3155	3188	rVB2	516454	917401	57.14%	7.684%
73	11.558	3259	3271	3303	rBV2	686179	1295950	80.72%	10.854%
74	11.992	3403	3406	3408	rBV	405	237	0.01%	0.002%
75	12.079	3431	3433	3436	rVB3	576	234	0.01%	0.002%
76	12.095	3436	3438	3441	rBV2	229	172	0.01%	0.001%

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041824\
 Data File : VV035292.D
 Acq On : 18 Apr 2024 18:50
 Operator : SY/MD
 Sample : P2157-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0RJ7

Integration Parameters: LSCINT.P
 Integrator: RTE
 Smoothing : OFF
 Sampling : 1
 Start Thrs: 0.2
 Stop Thrs : 0
 Filtering: 5
 Min Area: 0 % of largest Peak
 Max Peaks: 100
 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
 Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Title : VOC Analysis

77	12.400	3531	3533	3534	rBV	392	123	0.01%	0.001%
78	12.751	3638	3642	3644	rBV	520	395	0.02%	0.003%
79	12.886	3682	3684	3687	rBV	300	194	0.01%	0.002%
80	12.953	3702	3705	3709	rBV2	558	490	0.03%	0.004%
81	13.008	3721	3722	3723	rBV2	307	110	0.01%	0.001%
82	13.078	3742	3744	3748	rBV2	563	403	0.03%	0.003%
83	13.197	3770	3781	3798	rBV	107901	175777	10.95%	1.472%
84	13.680	3923	3931	3948	rVB2	28554	46440	2.89%	0.389%
85	15.336	4438	4446	4456	rBV3	17781	28225	1.76%	0.236%

Sum of corrected areas: 11939323

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041824\
Data File : VV035292.D
Acq On : 18 Apr 2024 18:50
Operator : SY/MD
Sample : P2157-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0RJ7

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041824\
Data File : VV035292.D
Acq On : 18 Apr 2024 18:50
Operator : SY/MD
Sample : P2157-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0RJ7

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc
------------------	----	---------	-------	----------	---	----	------	------