

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041922\
 Data File : VV025623.D
 Acq On : 19 Apr 2022 10:08
 Operator : SY/MD
 Sample : VV0419WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
ClientSampleId :
 VBLK347

Quant Time: Apr 20 02:50:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 20 02:46:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	193775	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	190546	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	87005	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	79718	4.425	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	88.400%	
7) Chloroethane-d5	1.568	69	71951	3.754	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	75.000%	
11) 1,1-Dichloroethene-d2	2.104	63	128609	3.564	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	71.200%	
20) 2-Butanone-d5	3.899	46	105150	47.983	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	95.960%	
24) Chloroform-d	4.346	84	123732	4.545	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.000%	
26) 1,2-Dichloroethane-d4	5.030	65	58060	4.950	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.000%	
32) Benzene-d6	5.046	84	251442	4.561	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.200%	
36) 1,2-Dichloropropane-d6	6.066	67	73590	4.786	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	95.800%	
41) Toluene-d8	7.313	98	214362	4.190	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	83.800%	
43) trans-1,3-Dichloroprop...	7.622	79	22709	4.394	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.800%	
46) 2-Hexanone-d5	8.088	63	78908	43.504	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	87.000%	
56) 1,1,2,2-Tetrachloroeth...	10.213	84	42145	4.339	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	86.800%	
66) 1,2-Dichlorobenzene-d4	11.622	152	69254	4.748	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	95.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041922\
 Data File : VV025623.D
 Acq On : 19 Apr 2022 10:08
 Operator : SY/MD
 Sample : VV0419WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK347

Quant Time: Apr 20 02:50:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 20 02:46:45 2022
 Response via : Initial Calibration

