

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041922\
 Data File : VW025633.D
 Acq On : 19 Apr 2022 16:47
 Operator : SY/MD
 Sample : N2384-20DL 20X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6S0DL

Quant Time: Apr 20 02:52:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 20 02:46:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	170169	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	176927	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	80860	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.311	65	60182	3.804	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.000%
7) Chloroethane-d5	1.571	69	60689	3.606	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.200%
11) 1,1-Dichloroethene-d2	2.111	63	93191	2.940	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.800%#
20) 2-Butanone-d5	3.896	46	90376	46.962	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.920%
24) Chloroform-d	4.352	84	103048	4.310	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.200%
26) 1,2-Dichloroethane-d4	5.034	65	49097	4.766	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.400%
32) Benzene-d6	5.053	84	202625	3.958	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.200%
36) 1,2-Dichloropropane-d6	6.069	67	62732	4.394	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.800%
41) Toluene-d8	7.317	98	171956	3.620	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.400%
43) trans-1,3-Dichloroprop...	7.625	79	18264	3.806	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.200%
46) 2-Hexanone-d5	8.088	63	66700	39.604	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.200%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	37521	4.160	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	58345	4.304	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.000%
Target Compounds						Qvalue
16) Methylene chloride	2.513	84	3104	0.217	ug/L	87
47) Tetrachloroethene	7.976	164	56559	5.206	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041922\
Data File : VV025633.D
Acq On : 19 Apr 2022 16:47
Operator : SY/MD
Sample : N2384-20DL 20X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6S0DL

Quant Time: Apr 20 02:52:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 20 02:46:45 2022
Response via : Initial Calibration

