

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041924\
 Data File : VW035305.D
 Acq On : 19 Apr 2024 14:08
 Operator : SY/MD
 Sample : P2157-13
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MCOB0

Quant Time: Apr 19 22:30:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 19 22:28:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	382890	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	370301	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	160887	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	119463	32.307	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	64.620%
7) Chloroethane-d5	1.532	69	108625	33.821	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	67.640%#
11) 1,1-Dichloroethene-d2	2.060	65	52166	32.183	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.360%
21) 2-Butanone-d5	3.789	46	196754	89.716	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.720%
24) Chloroform-d	4.249	84	261719	42.096	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.200%
26) 1,2-Dichloroethane-d4	4.941	65	169444	42.927	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.860%
32) Benzene-d6	4.960	84	495415	39.829	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.660%
36) 1,2-Dichloropropane-d6	5.985	67	171644	43.456	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.920%
41) Toluene-d8	7.243	98	402578	35.938	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	71.880%#
43) trans-1,3-Dichloroprop...	7.555	79	62728	32.667	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.340%
47) 2-Hexanone-d5	8.024	63	110561	74.034	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.030%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	249165	45.809	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.620%
66) 1,2-Dichlorobenzene-d4	11.558	152	164608	45.569	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041924\
Data File : VV035305.D
Acq On : 19 Apr 2024 14:08
Operator : SY/MD
Sample : P2157-13
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MCOB0

Quant Time: Apr 19 22:30:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 19 22:28:16 2024
Response via : Initial Calibration

