

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041924\
 Data File : VW035313.D
 Acq On : 19 Apr 2024 17:14
 Operator : SY/MD
 Sample : P2157-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CORJ8

Quant Time: Apr 19 22:32:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 19 22:28:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	405953	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	406452	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	181217	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	125346	31.972	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	63.940%	
7) Chloroethane-d5	1.532	69	113166	33.233	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	66.460%#	
11) 1,1-Dichloroethene-d2	2.056	65	56479	32.864	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	65.720%	
21) 2-Butanone-d5	3.790	46	192961	82.988	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	82.990%	
24) Chloroform-d	4.246	84	272241	41.301	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	82.600%	
26) 1,2-Dichloroethane-d4	4.941	65	177390	42.387	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	84.780%	
32) Benzene-d6	4.960	84	531103	38.900	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	77.800%	
36) 1,2-Dichloropropane-d6	5.986	67	179996	41.517	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	83.040%	
41) Toluene-d8	7.243	98	450354	36.627	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	73.260%#	
43) trans-1,3-Dichloroprop...	7.555	79	66218	31.418	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	62.840%	
47) 2-Hexanone-d5	8.024	63	109167	66.599	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	66.600%	
56) 1,1,2,2-Tetrachloroeth...	10.149	84	258685	43.329	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	86.660%	
66) 1,2-Dichlorobenzene-d4	11.558	152	184251	45.285	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	90.560%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041924\
Data File : VV035313.D
Acq On : 19 Apr 2024 17:14
Operator : SY/MD
Sample : P2157-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CORJ8

Quant Time: Apr 19 22:32:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 19 22:28:16 2024
Response via : Initial Calibration

