

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042022\
 Data File : VW025653.D
 Acq On : 20 Apr 2022 17:57
 Operator : SY/MD
 Sample : N2492-10
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6W2

Quant Time: Apr 21 06:44:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 21 06:38:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	136805	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	139397	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	58919	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	48827	3.839	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	76.800%	
7) Chloroethane-d5	1.574	69	49561	3.663	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	73.200%	
11) 1,1-Dichloroethene-d2	2.114	63	74509	2.924	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	58.400%#	
20) 2-Butanone-d5	3.899	46	90740	58.651	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	117.300%	
24) Chloroform-d	4.352	84	95240	4.955	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.200%	
26) 1,2-Dichloroethane-d4	5.037	65	45796	5.530	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	110.600%	
32) Benzene-d6	5.053	84	183836	4.558	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.200%	
36) 1,2-Dichloropropane-d6	6.072	67	58630	5.212	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	104.200%	
41) Toluene-d8	7.317	98	153980	4.114	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.200%	
43) trans-1,3-Dichloroprop...	7.625	79	17077	4.517	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	90.400%	
46) 2-Hexanone-d5	8.088	63	66093	49.810	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.620%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	36241	5.100	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	102.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	51716	5.235	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	104.800%	
Target Compounds						Qvalue
42) Toluene	7.400	91	10179	0.227	ug/L	96
47) Tetrachloroethene	7.976	164	34992	4.088	ug/L	96
52) Ethylbenzene	9.024	91	5008	0.105	ug/L	96
53) m,p-Xylene	9.146	106	1306	0.072	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042022\
Data File : VV025653.D
Acq On : 20 Apr 2022 17:57
Operator : SY/MD
Sample : N2492-10
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6W2

Quant Time: Apr 21 06:44:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 21 06:38:14 2022
Response via : Initial Calibration

