

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042022\
 Data File : VW025644.D
 Acq On : 20 Apr 2022 14:27
 Operator : SY/MD
 Sample : N2492-02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6T4

Quant Time: Apr 21 06:42:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 21 06:38:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	130157	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	134938	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	59743	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	49355	4.079	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.600%
7) Chloroethane-d5	1.571	69	50659	3.935	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.800%
11) 1,1-Dichloroethene-d2	2.111	63	78773	3.250	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.000%
20) 2-Butanone-d5	3.895	46	76820	52.189	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.380%
24) Chloroform-d	4.352	84	87754	4.799	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.000%
26) 1,2-Dichloroethane-d4	5.037	65	42048	5.337	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.800%
32) Benzene-d6	5.053	84	169406	4.339	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
36) 1,2-Dichloropropane-d6	6.069	67	53116	4.878	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.600%
41) Toluene-d8	7.317	98	141969	3.918	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.400%
43) trans-1,3-Dichloroprop...	7.625	79	14833	4.053	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.000%
46) 2-Hexanone-d5	8.088	63	55283	43.040	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.080%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	31472	4.575	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	49292	4.921	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042022\
Data File : VV025644.D
Acq On : 20 Apr 2022 14:27
Operator : SY/MD
Sample : N2492-02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6T4

Quant Time: Apr 21 06:42:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 21 06:38:14 2022
Response via : Initial Calibration

