

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042122\
 Data File : VW025661.D
 Acq On : 21 Apr 2022 13:04
 Operator : SY/MD
 Sample : N2471-04
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFFG6

Quant Time: Apr 22 03:07:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 22 03:06:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	140278	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	142648	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62683	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	49889	3.825	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.600%
7) Chloroethane-d5	1.571	69	49872	3.595	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	71.800%
11) 1,1-Dichloroethene-d2	2.111	63	76059	2.911	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.200%#
20) 2-Butanone-d5	3.899	46	84892	53.512	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.020%
24) Chloroform-d	4.349	84	92480	4.693	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.800%
26) 1,2-Dichloroethane-d4	5.034	65	44285	5.215	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
32) Benzene-d6	5.050	84	174777	4.235	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.600%
36) 1,2-Dichloropropane-d6	6.069	67	55659	4.835	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.800%
41) Toluene-d8	7.313	98	143469	3.746	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.000%
43) trans-1,3-Dichloroprop...	7.625	79	16268	4.205	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.000%
46) 2-Hexanone-d5	8.088	63	57794	42.563	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.120%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	32067	4.410	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	49895	4.748	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.000%
Target Compounds						Qvalue
16) Methylene chloride	2.510	84	2851	0.242	ug/L	95
25) Chloroform	4.378	83	9141	0.501	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042122\
Data File : VV025661.D
Acq On : 21 Apr 2022 13:04
Operator : SY/MD
Sample : N2471-04
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFFG6

Quant Time: Apr 22 03:07:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 22 03:06:52 2022
Response via : Initial Calibration

