

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042219\
 Data File : VV010427.D
 Acq On : 22 Apr 2019 22:34
 Operator : SY/MD
 Sample : K2456-15
 Misc : 5.10G/10mL/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAHN5

Quant Time: Apr 23 07:42:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM042219S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 23 06:57:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	467266	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	457928	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	219750	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	78176	12.14	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	48.56%
7) Chloroethane-d5	1.58	69	74281	14.83	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	59.32%
10) 1,1-Dichloroethene-d2	2.13	63	129354	10.76	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	43.04%#
20) 2-Butanone-d5	3.93	46	46385	30.81	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.62%
24) Chloroform-d	4.40	84	187983	16.46	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	65.84%
26) 1,2-Dichloroethane-d4	5.08	65	121948	17.41	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	69.64%#
29) Benzene-d6	5.10	84	362454	17.14	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	68.56%
33) 1,2-Dichloropropane-d6	6.12	67	111204	17.83	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	71.32%
37) Toluene-d8	7.36	98	351628	16.15	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	64.60%
38) trans-1,3-Dichloropropene-	7.66	79	45586	14.84	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	59.36%
39) 2-Hexanone-d5	8.13	63	35488	28.65	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	57.30%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	93596	15.40	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	61.60%
61) 1,2-Dichlorobenzene-d4	11.67	152	146717	17.57	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	70.28%#

Target Compounds					Ovalue
13) Acetone	2.19	43	4390	2.721 ug/L	79
16) Methylene chloride	2.53	84	10146	1.481 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042219\
Data File : VV010427.D
Acq On : 22 Apr 2019 22:34
Operator : SY/MD
Sample : K2456-15
Misc : 5.10G/10mL/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAHN5

Quant Time: Apr 23 07:42:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM042219S.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 23 06:57:14 2019
Response via : Initial Calibration

