

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042222\
 Data File : VW025686.D
 Acq On : 22 Apr 2022 17:29
 Operator : SY/MD
 Sample : N2492-09DL 200X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6W1DL

Quant Time: Apr 23 02:00:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 23 01:57:52 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	131318	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	134017	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	59908	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.311	65	75516	6.185	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 123.800%		
7) Chloroethane-d5	1.571	69	69438	5.346	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 107.000%		
11) 1,1-Dichloroethene-d2	2.111	63	112936	4.618	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 92.400%		
20) 2-Butanone-d5	3.899	46	95286	64.162	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 128.320%		
24) Chloroform-d	4.352	84	112251	6.084	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 121.600%		
26) 1,2-Dichloroethane-d4	5.037	65	52316	6.581	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 131.600%#		
32) Benzene-d6	5.053	84	225253	5.809	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 116.200%		
36) 1,2-Dichloropropane-d6	6.069	67	65976	6.101	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 122.000%		
41) Toluene-d8	7.317	98	184536	5.128	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 102.600%		
43) trans-1,3-Dichloroprop...	7.625	79	19393	5.335	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 106.800%		
46) 2-Hexanone-d5	8.088	63	69184	54.232	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 108.460%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	36850	5.394	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 107.800%		
66) 1,2-Dichlorobenzene-d4	11.622	152	56895	5.664	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 113.200%		
Target Compounds						
16) Methylene chloride	2.513	84	2576	0.233	ug/L	85
47) Tetrachloroethene	7.976	164	34389	4.179	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042222\
Data File : VV025686.D
Acq On : 22 Apr 2022 17:29
Operator : SY/MD
Sample : N2492-09DL 200X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6W1DL

Quant Time: Apr 23 02:00:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 23 01:57:52 2022
Response via : Initial Calibration

