

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042424\
 Data File : VW035384.D
 Acq On : 24 Apr 2024 14:01
 Operator : SY/MD
 Sample : P2238-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1CZ2

Quant Time: Apr 25 02:00:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 25 01:57:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	316782	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	323989	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	148997	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	151316	49.461	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.920%
7) Chloroethane-d5	1.532	69	129312	48.664	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.320%
11) 1,1-Dichloroethene-d2	2.060	65	64959	48.438	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.880%
21) 2-Butanone-d5	3.796	46	180364	99.405	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.410%
24) Chloroform-d	4.246	84	288362	56.061	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.120%
26) 1,2-Dichloroethane-d4	4.940	65	181877	55.693	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.380%
32) Benzene-d6	4.960	84	563705	51.797	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.600%
36) 1,2-Dichloropropane-d6	5.985	67	181636	52.559	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.120%
41) Toluene-d8	7.243	98	482411	49.220	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.440%
43) trans-1,3-Dichloroprop...	7.554	79	77384	46.060	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.120%
47) 2-Hexanone-d5	8.027	63	99888	76.448	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.450%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	263001	55.264	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	110.520%
66) 1,2-Dichlorobenzene-d4	11.558	152	199751	59.711	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	119.420%
Target Compounds						
13) Acetone	2.130	43	9499	5.702	ug/L	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042424\
Data File : VV035384.D
Acq On : 24 Apr 2024 14:01
Operator : SY/MD
Sample : P2238-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1CZ2

Quant Time: Apr 25 02:00:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 25 01:57:41 2024
Response via : Initial Calibration

