

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042522\
 Data File : VW025705.D
 Acq On : 25 Apr 2022 10:39
 Operator : SY/MD
 Sample : N2530-06
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6Z0

Quant Time: Apr 26 06:33:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 26 06:32:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	177608	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	179195	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	79099	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	95014	5.754	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	115.000%
7) Chloroethane-d5	1.571	69	94701	5.391	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.800%
11) 1,1-Dichloroethene-d2	2.111	63	148560	4.491	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.800%
20) 2-Butanone-d5	3.883	46	119309	59.400	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.800%
24) Chloroform-d	4.346	84	146498	5.871	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	117.400%
26) 1,2-Dichloroethane-d4	5.030	65	67461	6.275	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	125.400%
32) Benzene-d6	5.047	84	303345	5.851	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	117.000%
36) 1,2-Dichloropropane-d6	6.066	67	87201	6.031	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	120.600%
41) Toluene-d8	7.313	98	258055	5.363	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.200%
43) trans-1,3-Dichloroprop...	7.622	79	25514	5.250	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	105.000%
46) 2-Hexanone-d5	8.085	63	87881	51.520	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.040%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	47028	5.148	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	82487	6.220	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	124.400%#
Target Compounds						Qvalue
16) Methylene chloride	2.510	84	7132	0.478	ug/L	98
42) Toluene	7.391	91	19845	0.345	ug/L	99
53) m,p-Xylene	9.146	106	2552	0.109	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042522\
Data File : VV025705.D
Acq On : 25 Apr 2022 10:39
Operator : SY/MD
Sample : N2530-06
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6Z0

Quant Time: Apr 26 06:33:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 26 06:32:17 2022
Response via : Initial Calibration

