

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042522\
 Data File : VW025707.D
 Acq On : 25 Apr 2022 11:29
 Operator : SY/MD
 Sample : N2530-02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6X1

Quant Time: Apr 26 06:33:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 26 06:32:17 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.616	114	155975	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.850	117	160893	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.249	152	72406	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.310	65	93871	6.473	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	129.400%	
7) Chloroethane-d5		1.571	69	91781	5.949	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.000%	
11) 1,1-Dichloroethene-d2		2.111	63	142278	4.898	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.000%	
20) 2-Butanone-d5		3.886	46	119419	67.701	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	135.400%#	
24) Chloroform-d		4.349	84	145782	6.653	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	133.000%#	
26) 1,2-Dichloroethane-d4		5.031	65	68119	7.215	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	144.200%#	
32) Benzene-d6		5.050	84	302793	6.504	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	130.000%#	
36) 1,2-Dichloropropane-d6		6.069	67	86644	6.674	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	133.400%	
41) Toluene-d8		7.313	98	259049	5.996	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	120.000%	
43) trans-1,3-Dichloroprop...		7.622	79	26371	6.043	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	120.800%	
46) 2-Hexanone-d5		8.085	63	90505	59.094	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	118.180%	
56) 1,1,2,2-Tetrachloroeth...		10.214	84	49305	6.011	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	120.200%#	
66) 1,2-Dichlorobenzene-d4		11.622	152	81822	6.740	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	134.800%#	
Target Compounds							
14) Carbon disulfide		2.298	76	3136	0.097	ug/L #	92
42) Toluene		7.394	91	10830	0.209	ug/L	95
47) Tetrachloroethene		7.979	164	2350	0.238	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042522\
Data File : VV025707.D
Acq On : 25 Apr 2022 11:29
Operator : SY/MD
Sample : N2530-02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6X1

Quant Time: Apr 26 06:33:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 26 06:32:17 2022
Response via : Initial Calibration

