

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042922\
 Data File : VW025739.D
 Acq On : 29 Apr 2022 12:38
 Operator : SY/MD
 Sample : N2615-05
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH0D2

Quant Time: May 02 01:24:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon May 02 01:22:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	148750	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	154918	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	69753	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	53053	3.836	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	76.800%	
7) Chloroethane-d5	1.571	69	56895	3.867	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	77.400%	
11) 1,1-Dichloroethene-d2	2.111	63	86120	3.109	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	62.200%	
20) 2-Butanone-d5	3.892	46	99091	58.905	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	117.820%	
24) Chloroform-d	4.352	84	109154	5.223	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	104.400%	
26) 1,2-Dichloroethane-d4	5.034	65	50668	5.627	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	112.600%	
32) Benzene-d6	5.053	84	205376	4.582	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.600%	
36) 1,2-Dichloropropane-d6	6.069	67	63442	5.075	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.600%	
41) Toluene-d8	7.317	98	170323	4.095	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.800%	
43) trans-1,3-Dichloroprop...	7.625	79	18280	4.351	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.000%	
46) 2-Hexanone-d5	8.088	63	75814	51.411	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	102.820%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	40363	5.111	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	102.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	60983	5.215	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	104.200%	
Target Compounds						
5) Vinyl chloride	1.314	62	10207	0.619	ug/L	95
22) cis-1,2-Dichloroethene	3.918	96	23174	2.036	ug/L	97
51) Chlorobenzene	8.882	112	3137	0.099	ug/L	95
53) m,p-Xylene	9.143	106	1707	0.085	ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042922\
Data File : VV025739.D
Acq On : 29 Apr 2022 12:38
Operator : SY/MD
Sample : N2615-05
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH0D2

Quant Time: May 02 01:24:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon May 02 01:22:31 2022
Response via : Initial Calibration

