

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042922\
 Data File : VW025743.D
 Acq On : 29 Apr 2022 14:12
 Operator : SY/MD
 Sample : N2612-04
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH099

Quant Time: May 02 01:25:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon May 02 01:22:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	140326	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	150263	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	67798	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	49263	3.776	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.600%
7) Chloroethane-d5	1.568	69	57560	4.147	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.000%
11) 1,1-Dichloroethene-d2	2.108	63	80535	3.082	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.600%
20) 2-Butanone-d5	3.886	46	99524	62.714	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.420%
24) Chloroform-d	4.349	84	103108	5.230	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.600%
26) 1,2-Dichloroethane-d4	5.034	65	48249	5.680	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.600%
32) Benzene-d6	5.050	84	193499	4.451	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.000%
36) 1,2-Dichloropropane-d6	6.069	67	61263	5.053	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.000%
41) Toluene-d8	7.317	98	157841	3.912	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.200%
43) trans-1,3-Dichloroprop...	7.625	79	17781	4.363	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.200%
46) 2-Hexanone-d5	8.088	63	69168	48.357	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.720%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	38015	4.963	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	57391	5.049	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042922\
Data File : VV025743.D
Acq On : 29 Apr 2022 14:12
Operator : SY/MD
Sample : N2612-04
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH099

Quant Time: May 02 01:25:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon May 02 01:22:31 2022
Response via : Initial Calibration

