

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042922\
 Data File : VW025746.D
 Acq On : 29 Apr 2022 15:23
 Operator : SY/MD
 Sample : N2612-07
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH0A5

Quant Time: May 02 01:25:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon May 02 01:22:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	164649	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	172303	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	78992	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.311	65	54529	3.562	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.200%
7) Chloroethane-d5	1.571	69	59343	3.644	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.800%
11) 1,1-Dichloroethene-d2	2.111	63	89686	2.925	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.400%#
20) 2-Butanone-d5	3.896	46	113041	60.709	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.420%
24) Chloroform-d	4.352	84	119076	5.148	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.000%
26) 1,2-Dichloroethane-d4	5.037	65	55546	5.573	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.400%
32) Benzene-d6	5.053	84	222236	4.458	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.200%
36) 1,2-Dichloropropane-d6	6.069	67	69695	5.013	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.200%
41) Toluene-d8	7.317	98	184200	3.981	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.600%
43) trans-1,3-Dichloroprop...	7.625	79	20363	4.357	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.200%
46) 2-Hexanone-d5	8.088	63	81979	49.983	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.960%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	43273	4.926	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	65718	4.962	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042922\
Data File : VV025746.D
Acq On : 29 Apr 2022 15:23
Operator : SY/MD
Sample : N2612-07
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH0A5

Quant Time: May 02 01:25:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon May 02 01:22:31 2022
Response via : Initial Calibration

