

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042922\
 Data File : VW025755.D
 Acq On : 29 Apr 2022 18:54
 Operator : SY/MD
 Sample : N2612-22
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH0D3

Quant Time: May 02 01:27:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon May 02 01:22:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	153782	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	159189	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	73565	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	48628	3.401	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	68.000%	
7) Chloroethane-d5	1.571	69	52666	3.463	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	69.200%	
11) 1,1-Dichloroethene-d2	2.111	63	81957	2.862	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	57.200%#	
20) 2-Butanone-d5	3.895	46	107353	61.728	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	123.460%	
24) Chloroform-d	4.352	84	106427	4.926	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.600%	
26) 1,2-Dichloroethane-d4	5.034	65	49458	5.313	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.200%	
32) Benzene-d6	5.053	84	196940	4.276	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	85.600%	
36) 1,2-Dichloropropane-d6	6.069	67	63785	4.966	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.400%	
41) Toluene-d8	7.317	98	164915	3.858	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	77.200%	
43) trans-1,3-Dichloroprop...	7.625	79	16270	3.768	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	75.400%	
46) 2-Hexanone-d5	8.088	63	76377	50.403	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	100.800%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	42699	5.261	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	105.200%	
66) 1,2-Dichlorobenzene-d4	11.625	152	59948	4.860	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	97.200%	
Target Compounds						
					Qvalue	
22) cis-1,2-Dichloroethene	3.915	96	34326	2.917	ug/L	99
34) Trichloroethene	5.924	95	6149	0.492	ug/L	89
47) Tetrachloroethene	7.976	164	43397	4.440	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042922\
Data File : VV025755.D
Acq On : 29 Apr 2022 18:54
Operator : SY/MD
Sample : N2612-22
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH0D3

Quant Time: May 02 01:27:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon May 02 01:22:31 2022
Response via : Initial Calibration

