

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042922\
 Data File : VW025740.D
 Acq On : 29 Apr 2022 13:02
 Operator : SY/MD
 Sample : N2612-01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH096

Quant Time: May 02 01:24:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon May 02 01:22:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	155014	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	163636	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	73297	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	55267	3.835	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	76.600%	
7) Chloroethane-d5	1.571	69	59688	3.893	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	77.800%	
11) 1,1-Dichloroethene-d2	2.111	63	86266	2.988	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	59.800%#	
20) 2-Butanone-d5	3.892	46	104075	59.368	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	118.740%	
24) Chloroform-d	4.352	84	110084	5.055	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.000%	
26) 1,2-Dichloroethane-d4	5.034	65	51408	5.479	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	109.600%	
32) Benzene-d6	5.053	84	209379	4.422	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.400%	
36) 1,2-Dichloropropane-d6	6.069	67	65557	4.965	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.200%	
41) Toluene-d8	7.317	98	173412	3.947	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	79.000%	
43) trans-1,3-Dichloroprop...	7.625	79	18836	4.244	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	84.800%	
46) 2-Hexanone-d5	8.088	63	74791	48.015	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	96.040%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	40583	4.865	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	61362	4.993	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042922\
Data File : VV025740.D
Acq On : 29 Apr 2022 13:02
Operator : SY/MD
Sample : N2612-01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH096

Quant Time: May 02 01:24:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon May 02 01:22:31 2022
Response via : Initial Calibration

