

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050124\
 Data File : VV035506.D
 Acq On : 01 May 2024 19:01
 Operator : SY/MD
 Sample : P2280-05
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: May 02 01:30:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050124WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 02 01:07:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	447598	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	444623	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	198844	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	161082	46.674	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.340%
7) Chloroethane-d5	1.529	69	137520	54.071	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	108.140%
11) 1,1-Dichloroethene-d2	2.056	65	70593	47.650	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.300%
21) 2-Butanone-d5	3.793	46	223376	95.902	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.900%
24) Chloroform-d	4.246	84	309875	48.950	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.900%
26) 1,2-Dichloroethane-d4	4.940	65	189796	50.910	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.820%
32) Benzene-d6	4.957	84	620362	49.569	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.140%
36) 1,2-Dichloropropane-d6	5.985	67	197996	50.740	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.480%
41) Toluene-d8	7.243	98	536498	47.096	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.200%
43) trans-1,3-Dichloroprop...	7.551	79	85983	45.210	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.420%
47) 2-Hexanone-d5	8.024	63	177251	98.847	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.850%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	290703	48.890	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.780%
66) 1,2-Dichlorobenzene-d4	11.558	152	209746	53.092	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.180%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050124\
Data File : VV035506.D
Acq On : 01 May 2024 19:01
Operator : SY/MD
Sample : P2280-05
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: May 02 01:30:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050124WMA.M
Quant Title : VOC Analysis
QLast Update : Thu May 02 01:07:15 2024
Response via : Initial Calibration

