

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050124\
 Data File : VV035494.D
 Acq On : 01 May 2024 14:20
 Operator : SY/MD
 Sample : P2269-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: May 02 01:27:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050124WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 02 01:07:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	488462	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	476427	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	207175	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	179729	47.720	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	95.440%		
7) Chloroethane-d5	1.526	69	133185	47.986	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	95.980%		
11) 1,1-Dichloroethene-d2	2.053	65	77207	47.755	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	95.500%		
21) 2-Butanone-d5	3.793	46	256649	100.969	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	100.970%		
24) Chloroform-d	4.243	84	333657	48.297	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	96.600%		
26) 1,2-Dichloroethane-d4	4.937	65	208426	51.231	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	102.460%		
32) Benzene-d6	4.957	84	678869	50.623	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	101.240%		
36) 1,2-Dichloropropane-d6	5.986	67	215856	51.624	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	103.240%		
41) Toluene-d8	7.240	98	590854	48.405	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	96.800%		
43) trans-1,3-Dichloroprop...	7.551	79	97599	47.892	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	95.780%		
47) 2-Hexanone-d5	8.024	63	203430	105.873	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	105.870%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	311951	48.962	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	97.920%		
66) 1,2-Dichlorobenzene-d4	11.558	152	224253	54.482	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	108.960%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050124\
Data File : VV035494.D
Acq On : 01 May 2024 14:20
Operator : SY/MD
Sample : P2269-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: May 02 01:27:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050124WMA.M
Quant Title : VOC Analysis
QLast Update : Thu May 02 01:07:15 2024
Response via : Initial Calibration

