

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050224\
 Data File : VV035514.D
 Acq On : 02 May 2024 11:45
 Operator : SY/MD
 Sample : P2321-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1CZ3

Quant Time: May 03 05:13:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050124WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 03 05:13:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	508736	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	509167	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	226630	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	171205	43.645	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	87.300%		
7) Chloroethane-d5	1.529	69	147561	51.047	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	102.100%		
11) 1,1-Dichloroethene-d2	2.057	65	74977	44.527	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	89.060%		
21) 2-Butanone-d5	3.790	46	239268	90.379	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	90.380%		
24) Chloroform-d	4.246	84	334404	46.476	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.960%		
26) 1,2-Dichloroethane-d4	4.937	65	205149	48.416	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	96.840%		
32) Benzene-d6	4.957	84	664705	46.379	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.760%		
36) 1,2-Dichloropropane-d6	5.986	67	206229	46.151	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	92.300%		
41) Toluene-d8	7.243	98	573285	43.945	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	87.900%		
43) trans-1,3-Dichloroprop...	7.551	79	94634	43.452	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	86.900%		
47) 2-Hexanone-d5	8.024	63	185055	90.117	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	90.120%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	303030	44.503	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	89.000%		
66) 1,2-Dichlorobenzene-d4	11.558	152	228722	50.798	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	101.600%		
Target Compounds					Qvalue	
13) Acetone	2.124	43	8928	3.643	ug/L	98
15) Methyl Acetate	2.381	43	19466	4.896	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050224\
Data File : VV035514.D
Acq On : 02 May 2024 11:45
Operator : SY/MD
Sample : P2321-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1CZ3

Quant Time: May 03 05:13:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050124WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 03 05:13:05 2024
Response via : Initial Calibration

