

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050224\
 Data File : VV035517.D
 Acq On : 02 May 2024 12:54
 Operator : SY/MD
 Sample : P2321-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: May 03 05:14:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050124WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 03 05:13:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	393208	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	409335	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	176605	50.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	134591	44.392	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.780%
7) Chloroethane-d5	1.529	69	120866	54.097	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	108.200%
11) 1,1-Dichloroethene-d2	2.057	65	59910	46.033	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.060%
21) 2-Butanone-d5	3.793	46	235569	115.126	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	115.130%
24) Chloroform-d	4.246	84	286759	51.564	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.120%
26) 1,2-Dichloroethane-d4	4.937	65	185853	56.749	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	113.500%
32) Benzene-d6	4.957	84	552826	47.981	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.960%
36) 1,2-Dichloropropane-d6	5.986	67	183117	50.973	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.940%
41) Toluene-d8	7.243	98	461625	44.016	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.040%
43) trans-1,3-Dichloroprop...	7.555	79	82784	47.281	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.560%
47) 2-Hexanone-d5	8.024	63	174449	105.671	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.670%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	287214	52.468	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.940%
66) 1,2-Dichlorobenzene-d4	11.558	152	188178	53.631	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050224\
Data File : VV035517.D
Acq On : 02 May 2024 12:54
Operator : SY/MD
Sample : P2321-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: May 03 05:14:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050124WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 03 05:13:05 2024
Response via : Initial Calibration

