

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV050718\
 Data File : VV005824.D
 Acq On : 07 May 2018 18:27
 Operator : SY/MD
 Sample : J2638-20
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: May 08 01:43:42 2018
 Quant Method : W:\HPCHEM1\MSVOA V\METHOD\SOMVTR050518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue May 08 01:22:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	238821	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	234709	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	96449	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	97820	5.39	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	107.80%
7) Chloroethane-d5	1.58	69	82317	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.20%
11) 1,1-Dichloroethene-d2	2.13	63	158256	3.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.20%
20) 2-Butanone-d5	3.96	46	221585	62.14	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.28%
24) Chloroform-d	4.41	84	171528	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.80%
26) 1,2-Dichloroethane-d4	5.09	65	107133	5.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.40%
32) Benzene-d6	5.11	84	349346	5.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.60%
36) 1,2-Dichloropropane-d6	6.12	67	106661	5.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.00%
41) Toluene-d8	7.37	98	307364	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
43) trans-1,3-Dichloropropene-	7.67	79	37734	4.89	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.80%
46) 2-Hexanone-d5	8.14	63	153662	50.95	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.90%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	68586	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	106652	5.32	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV050718\
Data File : VV005824.D
Acq On : 07 May 2018 18:27
Operator : SY/MD
Sample : J2638-20
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: May 08 01:43:42 2018
Quant Method : W:\HPCHEM1\MSVOA V\METHOD\SOMVTR050518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue May 08 01:22:26 2018
Response via : Initial Calibration

