

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV050718\
 Data File : VV005826.D
 Acq On : 07 May 2018 19:22
 Operator : SY/MD
 Sample : J2636-11
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F1C16

Quant Time: May 08 01:47:22 2018
 Quant Method : W:\HPCHEM1\MSVOA V\METHOD\SOMVTR050518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue May 08 01:22:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	236068	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	243060	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	98938	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	87993	4.91	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.20%
7) Chloroethane-d5	1.58	69	78029	5.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.00%
11) 1,1-Dichloroethene-d2	2.13	63	143812	3.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.80%
20) 2-Butanone-d5	3.96	46	153918	43.67	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.34%
24) Chloroform-d	4.41	84	163501	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
26) 1,2-Dichloroethane-d4	5.09	65	99500	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.80%
32) Benzene-d6	5.11	84	311235	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
36) 1,2-Dichloropropane-d6	6.13	67	95366	4.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.20%
41) Toluene-d8	7.37	98	274973	4.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.40%
43) trans-1,3-Dichloropropene-	7.67	79	34780	4.35	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.00%
46) 2-Hexanone-d5	8.14	63	112356	35.98	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	71.96%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	61398	4.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	102187	4.97	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV050718\
Data File : VV005826.D
Acq On : 07 May 2018 19:22
Operator : SY/MD
Sample : J2636-11
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F1C16

Quant Time: May 08 01:47:22 2018
Quant Method : W:\HPCHEM1\MSVOA V\METHOD\SOMVTR050518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue May 08 01:22:26 2018
Response via : Initial Calibration

