

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV050718\
 Data File : VV005822.D
 Acq On : 07 May 2018 17:32
 Operator : SY/MD
 Sample : J2594-21
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: May 08 10:40:10 2018
 Quant Method : W:\HPCHEM1\MSVOA V\METHOD\SOMVTR050518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue May 08 01:22:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	231734	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	232497	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	96125	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	94258	5.36	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	107.20%
7) Chloroethane-d5	1.58	69	80557	5.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.20%
11) 1,1-Dichloroethene-d2	2.13	63	157409	4.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.20%
20) 2-Butanone-d5	3.96	46	211385	61.09	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	122.18%
24) Chloroform-d	4.41	84	163428	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.00%
26) 1,2-Dichloroethane-d4	5.09	65	103836	5.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.20%
32) Benzene-d6	5.11	84	330210	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.60%
36) 1,2-Dichloropropane-d6	6.13	67	102577	5.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.80%
41) Toluene-d8	7.37	98	301686	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.00%
43) trans-1,3-Dichloropropene-	7.67	79	38112	4.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.60%
46) 2-Hexanone-d5	8.14	63	151870	50.84	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.68%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	69258	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	108376	5.43	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA V\DATA\VV050718\
Data File : VV005822.D
Acq On : 07 May 2018 17:32
Operator : SY/MD
Sample : J2594-21
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: May 08 10:40:10 2018
Quant Method : W:\HPCHEM1\MSVOA V\METHOD\SOMVTR050518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue May 08 01:22:26 2018
Response via : Initial Calibration

