

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050919\
 Data File : VV010767.D
 Acq On : 09 May 2019 14:47
 Operator : SY/MD
 Sample : K2693-13
 Misc : 4.42G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFH64

Quant Time: May 10 06:36:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 10 05:43:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	58628	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	38976	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	7663	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	26078	28.63	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	114.52%
7) Chloroethane-d5	1.57	69	27516	36.71	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	146.84%
10) 1,1-Dichloroethene-d2	2.12	63	41884	21.20	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	84.80%
20) 2-Butanone-d5	3.94	46	14990	54.85	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.70%
24) Chloroform-d	4.40	84	62687	39.22	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	156.88%#
26) 1,2-Dichloroethane-d4	5.08	65	36683	36.75	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	147.00%#
29) Benzene-d6	5.10	84	90900	44.31	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	177.24%#
33) 1,2-Dichloropropane-d6	6.12	67	38845	59.42	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	237.68%#
37) Toluene-d8	7.36	98	68486	34.86	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	139.44%#
38) trans-1,3-Dichloropropene-	7.66	79	7721	25.91	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	103.64%
39) 2-Hexanone-d5	8.13	63	8558	62.56	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	125.12%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	25254	41.11	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	164.44%#
61) 1,2-Dichlorobenzene-d4	11.67	152	9369	31.57	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	126.28%#

Target Compounds					Ovalue
13) Acetone	2.19	43	4380	10.352 ug/L	98
16) Methylene chloride	2.53	84	3242	3.576 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050919\
Data File : VV010767.D
Acq On : 09 May 2019 14:47
Operator : SY/MD
Sample : K2693-13
Misc : 4.42G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFH64

Quant Time: May 10 06:36:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
Quant Title : VOC Analysis
QLast Update : Fri May 10 05:43:08 2019
Response via : Initial Calibration

