

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050919\
 Data File : VV010775.D
 Acq On : 09 May 2019 18:58
 Operator : SY/MD
 Sample : K2693-20
 Misc : 5.71G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFHF5

Quant Time: May 10 06:57:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 10 05:43:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	78489	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	73374	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	28958	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	22794	18.69	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.76%
7) Chloroethane-d5	1.57	69	21043	20.97	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.88%
10) 1,1-Dichloroethene-d2	2.12	63	38653	14.61	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.44%
20) 2-Butanone-d5	3.93	46	14939	40.83	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.66%
24) Chloroform-d	4.40	84	46254	21.62	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.48%
26) 1,2-Dichloroethane-d4	5.08	65	30037	22.48	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.92%
29) Benzene-d6	5.10	84	89591	23.20	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.80%
33) 1,2-Dichloropropane-d6	6.12	67	29471	23.95	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.80%
37) Toluene-d8	7.36	98	78683	21.27	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.08%
38) trans-1,3-Dichloropropene-	7.66	79	10891	19.41	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	77.64%
39) 2-Hexanone-d5	8.13	63	10278	39.91	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.82%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	25171	21.76	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.04%
61) 1,2-Dichlorobenzene-d4	11.67	152	25249	22.51	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.04%

Target Compounds					Ovalue
13) Acetone	2.19	43	1476	2.606 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050919\
Data File : VV010775.D
Acq On : 09 May 2019 18:58
Operator : SY/MD
Sample : K2693-20
Misc : 5.71G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFHF5

Quant Time: May 10 06:57:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
Quant Title : VOC Analysis
QLast Update : Fri May 10 05:43:08 2019
Response via : Initial Calibration

