

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050919\
 Data File : VV010809.D
 Acq On : 10 May 2019 12:22
 Operator : SY/MD
 Sample : K2613-04
 Misc : 5.00G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: May 10 23:40:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 10 22:59:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	85535	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	83907	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	35649	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	31919	24.02	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.08%
7) Chloroethane-d5	1.57	69	29136	26.65	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.60%
10) 1,1-Dichloroethene-d2	2.12	63	53585	18.59	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	74.36%
20) 2-Butanone-d5	3.93	46	16826	42.20	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.40%
24) Chloroform-d	4.40	84	54341	23.30	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.20%
26) 1,2-Dichloroethane-d4	5.08	65	35677	24.50	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.00%
29) Benzene-d6	5.10	84	114063	25.83	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	103.32%
33) 1,2-Dichloropropane-d6	6.12	67	36818	26.16	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.64%
37) Toluene-d8	7.36	98	104751	24.77	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	99.08%
38) trans-1,3-Dichloropropene-	7.66	79	13534	21.10	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.40%
39) 2-Hexanone-d5	8.13	63	12744	43.28	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.56%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	29430	22.25	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.00%
61) 1,2-Dichlorobenzene-d4	11.67	152	35509	25.72	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.88%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050919\
Data File : VV010809.D
Acq On : 10 May 2019 12:22
Operator : SY/MD
Sample : K2613-04
Misc : 5.00G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: May 10 23:40:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM050719S.M
Quant Title : VOC Analysis
QLast Update : Fri May 10 22:59:03 2019
Response via : Initial Calibration

