

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051124\
 Data File : VV035583.D
 Acq On : 10 May 2024 15:52
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV212

Quant Time: May 11 01:14:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051024.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 11 01:09:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	9778	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9753	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	5388	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	3345	0.500	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 100.000%	
4) 1,2-Dichloroethane-d4	4.969	65	4015	0.525	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 104.000%	
7) 1,2-Dichloropropane-d6	5.995	67	4401	0.515	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 104.000%	
8) Toluene-d8	7.253	98	10913	0.503	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 100.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3720	0.505	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 102.000%	
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	4513	0.489	ug/L	98
6) Trichloroethene	5.852	95	3749	0.501	ug/L	99
9) 1,2-Dibromoethane	8.305	107	2850	0.517	ug/L	98
12) 1,2,3-Trichloropropane	10.197	75	2852	0.512	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.356	75	646	0.490	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051124\
Data File : VV035583.D
Acq On : 10 May 2024 15:52
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV212

Quant Time: May 11 01:14:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051024.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 11 01:09:31 2024
Response via : Initial Calibration

