

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051124\
 Data File : VV035586.D
 Acq On : 10 May 2024 17:52
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5289

Quant Time: May 11 01:18:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051024.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 11 01:09:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	9211	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9325	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	5095	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	3351	0.532	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 106.000%	
4) 1,2-Dichloroethane-d4	4.969	65	3941	0.547	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 110.000%	
7) 1,2-Dichloropropane-d6	5.995	67	4448	0.545	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 108.000%	
8) Toluene-d8	7.253	98	10505	0.506	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 102.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3805	0.540	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 108.000%	
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	4486	0.516	ug/L	100
6) Trichloroethene	5.852	95	3566	0.498	ug/L	99
9) 1,2-Dibromoethane	8.305	107	2778	0.527	ug/L	98
12) 1,2,3-Trichloropropane	10.197	75	2764	0.524	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.356	75	597	0.479	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051124\
Data File : VV035586.D
Acq On : 10 May 2024 17:52
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5289

Quant Time: May 11 01:18:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051024.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 11 01:09:31 2024
Response via : Initial Calibration

