

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051419\
Quantitation Report (QT Reviewed)


Data File : VV010862.D
Acq On : 14 May 2019 15:10
Operator : SY/MD
Sample : K2737-12
Misc : 3.76G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB860

Manual Integrations
APPROVED

MMDadoda
5/20/2019 1:04:17 AM

Quant Time: May 15 07:46:52 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM051319S.M
Quant Title : VOC Analysis
QLast Update : Wed May 15 06:52:21 2019
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051419\
Quantitation Report (Qedit)




Data File : VV010862.D
Acq On : 14 May 2019 15:10
Operator : SY/MD
Sample : K2737-12
Misc : 3.76G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB860

Manual Integrations
APPROVED

MMDadoda
5/20/2019 1:04:17 AM

Quant Time: May 15 06:56:27 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM051319S.M
Quant Title : VOC Analysis
QLast Update : Wed May 15 06:52:21 2019
Response via : Initial Calibration



TIC: VV010862.D

(13) Acetone (T)

2.196min (+0.007) 123.18ug/L

response 59201

Ion	Exp%	Act%
43.00	100	100
58.00	34.60	32.69
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051419\
Quantitation Report (Qedit)




Data File : VV010862.D
Acq On : 14 May 2019 15:10
Operator : SY/MD
Sample : K2737-12
Misc : 3.76G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB860

Manual Integrations
APPROVED

MMDadoda
5/20/2019 1:04:17 AM

Quant Time: May 15 06:56:27 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM051319S.M
Quant Title : VOC Analysis
QLast Update : Wed May 15 06:52:21 2019
Response via : Initial Calibration



(13) Acetone (T)

2.196min (+0.007) 283.89ug/L m

205/20/19 Sy

response 136437

Ion	Exp%	Act%
43.00	100	100
58.00	34.60	14.18
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051419\
Quantitation Report (QT Reviewed)

Data File : VV010862.D
Acq On : 14 May 2019 15:10
Operator : SY/MD
Sample : K2737-12
Misc : 3.76G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB860

Manual Integrations
APPROVED

MMDadoda
5/20/2019 1:04:17 AM

Quant Time: May 15 07:46:52 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM051319S.M
Quant Title : VOC Analysis
QLast Update : Wed May 15 06:52:21 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	117713	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	126037	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	53348	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	27753	18.51	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.04%
7) Chloroethane-d5	1.57	69	22917	23.45	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.80%
10) 1,1-Dichloroethene-d2	2.12	63	42122	14.84	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.36%
20) 2-Butanone-d5	3.94	46	20801	44.09	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.18%
24) Chloroform-d	4.40	84	70479	20.74	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.96%
26) 1,2-Dichloroethane-d4	5.08	65	44906	23.24	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.96%
29) Benzene-d6	5.10	84	151705	21.75	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.00%
33) 1,2-Dichloropropane-d6	6.12	67	50833	22.84	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	91.36%
37) Toluene-d8	7.36	98	135248	20.18	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	80.72%
38) trans-1,3-Dichloropropene-	7.66	79	16908	17.41	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	69.64%
39) 2-Hexanone-d5	8.13	63	14640	34.71	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	69.42%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	40067	19.41	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	77.64%
61) 1,2-Dichlorobenzene-d4	11.67	152	52185	23.42	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.68%

Target Compounds

13) Acetone	2.20	43	136437m)	283.894	ug/L	Qvalue
16) Methylene chloride	2.53	84	5264	2.771	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed