

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051424\
 Data File : VV035596.D
 Acq On : 13 May 2024 13:33
 Operator : SY/MD
 Sample : VV0514WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK206

Quant Time: May 14 02:06:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051024.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 11 01:09:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	10527	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	10552	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	5839	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	3515	0.488	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	98.000%
4) 1,2-Dichloroethane-d4	4.969	65	4178	0.507	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.000%
7) 1,2-Dichloropropane-d6	5.995	67	4606	0.499	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	100.000%
8) Toluene-d8	7.253	98	11555	0.492	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3944	0.495	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	98.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051424\
Data File : VV035596.D
Acq On : 13 May 2024 13:33
Operator : SY/MD
Sample : VV0514WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK206

Quant Time: May 14 02:06:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051024.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 11 01:09:31 2024
Response via : Initial Calibration

