




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051524\
Data File : VV035607.D
Acq On : 14 May 2024 13:20
Operator : SY/MD
Sample : P2416-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHAS3

Manual IntegrationsAPPROVED

Quant Time: May 15 02:39:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR051424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 15 02:29:38 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 05/15/2024
Supervised By :Semsettin Yesilyurt 05/15/2024



TIC: VV035607.D\data.ms

(13) Acetone (T)

2.150min (-0.003) 1.36 ug/L

response 3687

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	34.60	63.41
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051524\
Data File : VV035607.D
Acq On : 14 May 2024 13:20
Operator : SY/MD
Sample : P2416-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHAS3

Manual IntegrationsAPPROVED

Quant Time: May 15 02:39:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR051424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 15 02:29:38 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 05/15/2024
Supervised By :Semsettin Yesilyurt 05/15/2024



TIC: VV035607.D\data.ms

(13) Acetone (T)

2.169min (+ 0.016) 5.84 ug/L m

response 15779

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	34.60	14.82
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051524\
 Data File : VV035607.D
 Acq On : 14 May 2024 13:20
 Operator : SY/MD
 Sample : P2416-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHAS3

Manual IntegrationsAPPROVED

Quant Time: May 15 02:39:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR051424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 15 02:29:38 2024
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 05/15/2024
 Supervised By :Semsettin Yesilyurt 05/15/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	230416	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	236807	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	105864	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	66412	4.544	ug/L	0.00
Spiked Amount 5.000	Range 40	- 130	Recovery	=	90.800%	
7) Chloroethane-d5	1.532	69	62554	4.727	ug/L	0.00
Spiked Amount 5.000	Range 65	- 130	Recovery	=	94.600%	
11) 1,1-Dichloroethene-d2	2.056	65	30316	4.669	ug/L	0.00
Spiked Amount 5.000	Range 60	- 125	Recovery	=	93.400%	
20) 2-Butanone-d5	3.825	46	158172	46.438	ug/L	0.00
Spiked Amount 50.000	Range 40	- 130	Recovery	=	92.880%	
24) Chloroform-d	4.246	84	143925	4.641	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125	Recovery	=	92.800%	
26) 1,2-Dichloroethane-d4	4.941	65	70352	4.947	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130	Recovery	=	99.000%	
32) Benzene-d6	4.957	84	270009	4.975	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125	Recovery	=	99.400%	
36) 1,2-Dichloropropane-d6	5.986	67	91226	5.304	ug/L	0.00
Spiked Amount 5.000	Range 60	- 140	Recovery	=	106.000%	
41) Toluene-d8	7.243	98	208385	4.354	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130	Recovery	=	87.000%	
43) trans-1,3-Dichloroprop...	7.558	79	27084	4.526	ug/L	0.00
Spiked Amount 5.000	Range 55	- 130	Recovery	=	90.600%	
46) 2-Hexanone-d5	8.027	63	119863	44.281	ug/L	0.00
Spiked Amount 50.000	Range 45	- 130	Recovery	=	88.560%	
56) 1,1,2,2-Tetrachloroeth...	10.153	84	56925	4.085	ug/L	0.00
Spiked Amount 5.000	Range 65	- 120	Recovery	=	81.800%	
66) 1,2-Dichlorobenzene-d4	11.558	152	86764	5.300	ug/L	0.00
Spiked Amount 5.000	Range 80	- 120	Recovery	=	106.000%	
Target Compounds						
					Qvalue	
13) Acetone	2.169	43	15779m	5.835	ug/L	
16) Methylene chloride	2.446	84	9312	0.407	ug/L	97
25) Chloroform	4.278	83	5741	0.164	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051524\
Data File : VV035607.D
Acq On : 14 May 2024 13:20
Operator : SY/MD
Sample : P2416-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHAS3

Manual IntegrationsAPPROVED

Quant Time: May 15 02:39:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR051424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 15 02:29:38 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 05/15/2024
Supervised By :Semsettin Yesilyurt 05/15/2024

