

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW051622\
 Data File : VW025909.D
 Acq On : 16 May 2022 13:57
 Operator : SY/MD
 Sample : N2831-03
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YB2K0

Quant Time: May 17 08:01:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR051322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 17 07:58:29 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	134040	5.000	ug/L	0.01
28) Chlorobenzene-d5	8.853	117	138833	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	60999	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	60363	5.453	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 109.000%		
7) Chloroethane-d5	1.568	69	53768	5.672	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 113.400%		
11) 1,1-Dichloroethene-d2	2.117	63	233513	9.530	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 190.600%#		
20) 2-Butanone-d5	3.899	46	82230	64.140	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 128.280%		
24) Chloroform-d	4.352	84	99879	5.523	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 110.400%		
26) 1,2-Dichloroethane-d4	5.037	65	43483	5.670	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 113.400%		
32) Benzene-d6	5.053	84	200480	5.561	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 111.200%		
36) 1,2-Dichloropropane-d6	6.072	67	59660	5.692	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 113.800%		
41) Toluene-d8	7.317	98	167728	5.005	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 100.000%		
43) trans-1,3-Dichloroprop...	7.628	79	15601	4.551	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 91.000%		
46) 2-Hexanone-d5	8.091	63	59415	46.685	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 93.360%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	37000	5.480	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 109.600%		
66) 1,2-Dichlorobenzene-d4	11.622	152	51755	5.959	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 119.200%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.121	96	244354	18.969	ug/L #	85
14) Carbon disulfide	2.297	76	11071	0.334	ug/L	97
18) trans-1,2-Dichloroethene	2.767	96	4611	0.410	ug/L	98
19) 1,1-Dichloroethane	3.195	63	34619	1.628	ug/L	96
22) cis-1,2-Dichloroethene	3.912	96	165557	144.256	ug/L	97
27) 1,2-Dichloroethane	5.143	62	7804	0.687	ug/L	98
33) Benzene	5.108	78	13382	0.285	ug/L	100
34) Trichloroethene	5.915	95	78258	6.517	ug/L	100
47) Tetrachloroethene	7.982	164	2885	0.310	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051622\
Data File : VV025909.D
Acq On : 16 May 2022 13:57
Operator : SY/MD
Sample : N2831-03
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YB2K0

Quant Time: May 17 08:01:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR051322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 17 07:58:29 2022
Response via : Initial Calibration

