

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051622\
 Data File : VV025912.D
 Acq On : 16 May 2022 15:08
 Operator : SY/MD
 Sample : N2831-06
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YB2K3

Quant Time: May 17 08:02:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR051322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 17 07:58:29 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	155877	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	162068	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	70854	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	67449	5.239	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 104.800%		
7) Chloroethane-d5	1.568	69	59052	5.357	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 107.200%		
11) 1,1-Dichloroethene-d2	2.108	63	115900	4.067	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 81.400%		
20) 2-Butanone-d5	3.918	46	80906	54.266	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 108.540%		
24) Chloroform-d	4.352	84	106246	5.052	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 101.000%		
26) 1,2-Dichloroethane-d4	5.037	65	49941	5.600	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 112.000%		
32) Benzene-d6	5.050	84	210693	5.006	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 100.200%		
36) 1,2-Dichloropropane-d6	6.072	67	64533	5.274	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 105.400%		
41) Toluene-d8	7.317	98	176303	4.506	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 90.200%		
43) trans-1,3-Dichloroprop...	7.629	79	14287	3.571	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 71.400%		
46) 2-Hexanone-d5	8.092	63	76520	51.505	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 103.000%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	42494	5.392	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 107.800%		
66) 1,2-Dichlorobenzene-d4	11.625	152	58217	5.771	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 115.400%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.121	96	19521	1.303	ug/L #	33
14) Carbon disulfide	2.294	76	8848	0.229	ug/L #	94
17) Methyl tert-butyl Ether	2.770	73	3461	0.145	ug/L #	90
19) 1,1-Dichloroethane	3.195	63	10485	0.424	ug/L	95
22) cis-1,2-Dichloroethene	3.915	96	42157	3.159	ug/L	95
34) Trichloroethene	5.915	95	116094	8.282	ug/L	98
47) Tetrachloroethene	7.976	164	65449	6.026	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051622\
Data File : VV025912.D
Acq On : 16 May 2022 15:08
Operator : SY/MD
Sample : N2831-06
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YB2K3

Quant Time: May 17 08:02:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR051322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 17 07:58:29 2022
Response via : Initial Calibration

