

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051622\
 Data File : VV025900.D
 Acq On : 16 May 2022 10:24
 Operator : SY/MD
 Sample : VV0516WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK359

Quant Time: May 17 07:59:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR051322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 17 07:58:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	189757	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	188261	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	86377	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	89348	5.701	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.000%
7) Chloroethane-d5	1.561	69	76883	5.729	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.600%
11) 1,1-Dichloroethene-d2	2.101	63	130445	3.760	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.200%
20) 2-Butanone-d5	3.912	46	93760	51.660	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.320%
24) Chloroform-d	4.342	84	128167	5.007	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
26) 1,2-Dichloroethane-d4	5.027	65	58123	5.354	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.000%
32) Benzene-d6	5.043	84	261555	5.350	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.000%
36) 1,2-Dichloropropane-d6	6.066	67	77400	5.446	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.000%
41) Toluene-d8	7.313	98	222345	4.893	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.800%
43) trans-1,3-Dichloroprop...	7.622	79	24221	5.211	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	104.200%
46) 2-Hexanone-d5	8.088	63	79015	45.785	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.560%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	44639	4.876	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	66005	5.367	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051622\
Data File : VV025900.D
Acq On : 16 May 2022 10:24
Operator : SY/MD
Sample : VV0516WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK359

Quant Time: May 17 07:59:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR051322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 17 07:58:29 2022
Response via : Initial Calibration

