

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051623\
 Data File : VV031059.D
 Acq On : 16 May 2023 17:39
 Operator : SY/MD
 Sample : 02768-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CONM7

Quant Time: May 17 01:46:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 17 01:41:23 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	117784	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	115992	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	60039	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	37877	3.903	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	78.000%	
7) Chloroethane-d5	1.539	69	39256	4.959	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	99.200%	
11) 1,1-Dichloroethene-d2	2.063	65	21056	4.100	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	82.000%	
20) 2-Butanone-d5	3.818	46	100883	58.624	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	117.240%	
24) Chloroform-d	4.262	84	96536	5.006	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.200%	
26) 1,2-Dichloroethane-d4	4.953	65	53681	5.475	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	109.600%	
32) Benzene-d6	4.969	84	158263	4.655	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.000%	
36) 1,2-Dichloropropane-d6	5.998	67	48118	4.867	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	97.400%	
41) Toluene-d8	7.252	98	131898	4.074	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.400%	
43) trans-1,3-Dichloroprop...	7.564	79	15946	4.303	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.000%	
46) 2-Hexanone-d5	8.034	63	71303	54.915	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	109.840%	
56) 1,1,2,2-Tetrachloroeth...	10.159	84	38101	5.317	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.400%	
66) 1,2-Dichlorobenzene-d4	11.567	152	55918	4.996	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.000%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.220	50	1511	0.138	ug/L #	76
6) Bromomethane	1.494	94	543	0.125	ug/L	74
35) Methylcyclohexane	5.998	83	13406	1.112	ug/L #	18
37) 1,2-Dichloropropane	5.998	63	6915	0.999	ug/L #	86
48) 2-Hexanone	8.088	43	7721	2.807	ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051623\
Data File : VV031059.D
Acq On : 16 May 2023 17:39
Operator : SY/MD
Sample : 02768-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CONM7

Quant Time: May 17 01:46:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 17 01:41:23 2023
Response via : Initial Calibration









#3
Chloromethane
Concen: 0.138 ug/L
RT: 1.220 min Scan# 50
Delta R.T. 0.003 min
Lab File: VV031059.D
Acq: 16 May 2023 17:39




Instrument :
MSVOA_V
ClientSampleId :
CONM7


Tgt Ion: 50 Resp: 1511
Ion Ratio Lower Upper
50 100
52 45.9 22.8 42.4#



#6
Bromomethane
Concen: 0.125 ug/L
RT: 1.494 min Scan# 141
Delta R.T. 0.003 min
Lab File: VV031059.D
Acq: 16 May 2023 17:39

Tgt Ion: 94 Resp: 543
Ion Ratio Lower Upper
94 100
96 65.7 63.4 117.8









#35
Methylcyclohexane
Concen: 1.112 ug/L
RT: 5.998 min Scan# 1542
Delta R.T. -0.058 min
Lab File: VV031059.D
Acq: 16 May 2023 17:39




Instrument :
MSVOA_V
ClientSampleId :
CONM7


Tgt Ion: 83 Resp: 13406
Ion Ratio Lower Upper
83 100
55 1.1 62.9 94.3#
98 0.2 40.5 60.7#



#37
1,2-Dichloropropane
Concen: 0.999 ug/L
RT: 5.998 min Scan# 1542
Delta R.T. -0.100 min
Lab File: VV031059.D
Acq: 16 May 2023 17:39


Tgt Ion: 63 Resp: 6915
Ion Ratio Lower Upper
63 100
112 0.3 3.8 5.8#





#48
 2-Hexanone
 Concen: 2.807 ug/L
 RT: 8.088 min Scan# 21
 Delta R.T. 0.003 min
 Lab File: VV031059.D
 Acq: 16 May 2023 17:39

Instrument :
 MSVOA_V
 ClientSampleId :
 CONM7



Tgt Ion: 43 Resp: 7721

Ion	Ratio	Lower	Upper
43	100		
58	54.6	39.9	59.9
57	15.4	14.1	21.1
100	11.0	9.7	14.5

