

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW051624\
 Data File : VW035625.D
 Acq On : 16 May 2024 13:08
 Operator : SY/MD
 Sample : P2484-09
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: May 17 01:57:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050124WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 17 01:55:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	460885	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	466999	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	191292	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	129718	36.503	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	73.000%	
7) Chloroethane-d5	1.532	69	124114	47.393	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	94.780%	
11) 1,1-Dichloroethene-d2	2.060	65	61654	40.416	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	80.840%	
21) 2-Butanone-d5	3.793	46	223139	93.038	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	93.040%	
24) Chloroform-d	4.246	84	305706	46.899	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	93.800%	
26) 1,2-Dichloroethane-d4	4.940	65	192900	50.251	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.500%	
32) Benzene-d6	4.957	84	582877	44.342	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	88.680%	
36) 1,2-Dichloropropane-d6	5.985	67	195878	47.792	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	95.580%	
41) Toluene-d8	7.239	98	482624	40.336	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	80.680%	
43) trans-1,3-Dichloroprop...	7.554	79	79654	39.876	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	79.760%	
47) 2-Hexanone-d5	8.024	63	149589	79.424	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	79.420%	
56) 1,1,2,2-Tetrachloroeth...	10.149	84	265232	42.469	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	84.940%	
66) 1,2-Dichlorobenzene-d4	11.558	152	193885	51.015	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	102.040%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051624\
Data File : VV035625.D
Acq On : 16 May 2024 13:08
Operator : SY/MD
Sample : P2484-09
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: May 17 01:57:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM050124WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 17 01:55:25 2024
Response via : Initial Calibration

