

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051922\
 Data File : VV025935.D
 Acq On : 19 May 2022 11:19
 Operator : SY/MD
 Sample : VV0519WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK201

Quant Time: May 20 06:16:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051822.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 19 07:42:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	18998	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	17291	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	10009	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	5390	0.404	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	80.000%
4) 1,2-Dichloroethane-d4	5.042	65	3128	0.398	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	80.000%
7) 1,2-Dichloropropane-d6	6.068	67	3821	0.409	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	82.000%
8) Toluene-d8	7.317	98	14602	0.407	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	82.000%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	2926	0.405	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	82.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051922\
Data File : VV025935.D
Acq On : 19 May 2022 11:19
Operator : SY/MD
Sample : VV0519WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK201

Quant Time: May 20 06:16:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051822.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 19 07:42:58 2022
Response via : Initial Calibration

