

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051922\
 Data File : VV025937.D
 Acq On : 19 May 2022 13:15
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5282

Quant Time: May 20 06:16:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051822.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 19 07:42:58 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	20279	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	18079	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	10403	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	7264	0.510	ug/L	0.00
Spiked Amount	0.500	Range 40 - 130	Recovery	=	102.000%	
4) 1,2-Dichloroethane-d4	5.042	65	4450	0.531	ug/L	0.00
Spiked Amount	0.500	Range 70 - 130	Recovery	=	106.000%	
7) 1,2-Dichloropropane-d6	6.068	67	5377	0.550	ug/L	0.00
Spiked Amount	0.500	Range 60 - 140	Recovery	=	110.000%	
8) Toluene-d8	7.317	98	20684	0.551	ug/L	0.00
Spiked Amount	0.500	Range 70 - 130	Recovery	=	110.000%	
10) 1,1,2,2-Tetrachloroeth...	10.216	84	4146	0.549	ug/L	0.00
Spiked Amount	0.500	Range 65 - 120	Recovery	=	110.000%	
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	6286	0.499	ug/L	88
6) Trichloroethene	5.925	95	6442	0.513	ug/L	99
9) 1,2-Dibromoethane	8.353	107	4098	0.533	ug/L #	97
12) 1,2,3-Trichloropropane	10.267	75	2525	0.541	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.426	75	457	0.553	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV051922\
Data File : VV025937.D
Acq On : 19 May 2022 13:15
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5282

Quant Time: May 20 06:16:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM051822.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 19 07:42:58 2022
Response via : Initial Calibration

