

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV052019\
 Data File : VV010962.D
 Acq On : 20 May 2019 16:37
 Operator : SY/MD
 Sample : K2788-15
 Misc : 6.51G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A51A6

Quant Time: May 21 04:18:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM052019S.M
 Quant Title : VOC Analysis
 QLast Update : Tue May 21 04:03:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	149664	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	147003	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	61855	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	37719	22.11	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.44%
7) Chloroethane-d5	1.57	69	24974	21.62	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.48%
10) 1,1-Dichloroethene-d2	2.12	63	51528	15.71	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.84%
20) 2-Butanone-d5	3.94	46	27557	46.81	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.62%
24) Chloroform-d	4.40	84	77565	20.01	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	80.04%
26) 1,2-Dichloroethane-d4	5.08	65	43712	19.13	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	76.52%
29) Benzene-d6	5.09	84	167213	20.84	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	83.36%
33) 1,2-Dichloropropane-d6	6.12	67	49997	19.61	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	78.44%
37) Toluene-d8	7.36	98	152090	20.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.56%
38) trans-1,3-Dichloropropene-	7.66	79	21225	20.27	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	81.08%
39) 2-Hexanone-d5	8.13	63	21460	45.12	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.24%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	50271	19.75	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	79.00%
61) 1,2-Dichlorobenzene-d4	11.67	152	49704	19.79	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	79.16%

Target Compounds

					Ovalue
13) Acetone	2.19	43	5955	10.489	ug/L 100
14) Carbon disulfide	2.31	76	18090	4.090	ug/L 100
22) cis-1,2-Dichloroethene	3.96	96	211373	87.132	ug/L 100
31) 1,1,1-Trichloroethane	4.65	97	20950	6.335	ug/L 99
35) Trichloroethene	5.96	95	5527112	2360.612	ug/L 98
47) Tetrachloroethene	8.02	164	308491	168.791	ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV052019\
Data File : VV010962.D
Acq On : 20 May 2019 16:37
Operator : SY/MD
Sample : K2788-15
Misc : 6.51G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A51A6

Quant Time: May 21 04:18:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM052019S.M
Quant Title : VOC Analysis
QLast Update : Tue May 21 04:03:21 2019
Response via : Initial Calibration

