

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060120\
 Data File : VV016400.D
 Acq On : 01 Jun 2020 10:24
 Operator : SY/MD
 Sample : L2826-04DL 4X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CWKB7DL

Quant Time: Jun 02 01:30:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 02 01:13:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	89221	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	83715	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	35714	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	33841	4.33	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.60%
7) Chloroethane-d5	1.58	69	26655	4.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.80%
11) 1,1-Dichloroethene-d2	2.12	63	47610	3.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.80%
20) 2-Butanone-d5	3.96	46	89234	52.95	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.90%
24) Chloroform-d	4.38	84	60839	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.00%
26) 1,2-Dichloroethane-d4	5.06	65	34512	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.40%
32) Benzene-d6	5.07	84	118274	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.40%
36) 1,2-Dichloropropane-d6	6.10	67	35788	4.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.40%
41) Toluene-d8	7.34	98	102153	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
45) trans-1,3-Dichloropropene-	7.64	79	12746	4.81	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.20%
48) 2-Hexanone-d5	8.12	63	65995	53.54	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.08%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	27234	5.25	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.00%
65) 1,2-Dichlorobenzene-d4	11.65	152	32609	4.92	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.40%

Target Compounds

					Ovalue
18) trans-1,2-Dichloroethene	2.78	96	5286	0.867 ug/L	98
22) cis-1,2-Dichloroethene	3.94	96	10365	1.541 ug/L	90
25) Chloroform	4.41	83	2550	0.212 ug/L	96
34) Trichloroethene	5.94	95	1953	0.301 ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060120\
Data File : VV016400.D
Acq On : 01 Jun 2020 10:24
Operator : SY/MD
Sample : L2826-04DL 4X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWKB7DL

Quant Time: Jun 02 01:30:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 02 01:13:11 2020
Response via : Initial Calibration

