

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060120\
 Data File : VV016404.D
 Acq On : 01 Jun 2020 12:01
 Operator : SY/MD
 Sample : L2826-01 10X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CWKB4

Quant Time: Jun 02 02:13:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 02 01:13:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	83801	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	76737	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	25628	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	27355	3.73	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.60%
7) Chloroethane-d5	1.58	69	22305	4.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.00%
11) 1,1-Dichloroethene-d2	2.12	63	39314	3.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.20%
20) 2-Butanone-d5	3.96	46	70318	44.42	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.84%
24) Chloroform-d	4.38	84	50626	4.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.60%
26) 1,2-Dichloroethane-d4	5.06	65	28677	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.80%
32) Benzene-d6	5.08	84	100649	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.40%
36) 1,2-Dichloropropane-d6	6.10	67	30055	4.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.40%
41) Toluene-d8	7.34	98	85144	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
45) trans-1,3-Dichloropropene-	7.64	79	9734	4.01	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.20%
48) 2-Hexanone-d5	8.12	63	53037	46.94	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.88%
59) 1,1,2,2-Tetrachloroethane-	10.24	84	19196	4.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.60%
65) 1,2-Dichlorobenzene-d4	11.65	152	19497	4.10	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	82.00%

Target Compounds					Ovalue
16) Methylene chloride	2.52	84	1176	0.204 ug/L	93
21) 2-Butanone	4.06	43	29417	18.736 ug/L #	64
25) Chloroform	4.40	83	3067	0.271 ug/L	88
34) Trichloroethene	5.94	95	1082	0.182 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060120\
Data File : VV016404.D
Acq On : 01 Jun 2020 12:01
Operator : SY/MD
Sample : L2826-01 10X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWKB4

Quant Time: Jun 02 02:13:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 02 01:13:11 2020
Response via : Initial Calibration

