

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060120\
 Data File : VV016411.D
 Acq On : 01 Jun 2020 14:54
 Operator : SY/MD
 Sample : L2780-16
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AL9

Quant Time: Jun 02 03:03:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 02 01:13:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	55338	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	52338	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.27	152	22625	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	27739	5.73	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.60%
7) Chloroethane-d5	1.58	69	21838	5.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	118.60%
11) 1,1-Dichloroethene-d2	2.12	63	47771	5.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	112.80%
20) 2-Butanone-d5	3.96	46	62182	59.49	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.98%
24) Chloroform-d	4.37	84	58440	7.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	142.60%#
26) 1,2-Dichloroethane-d4	5.06	65	31669	7.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	148.40%#
32) Benzene-d6	5.07	84	111914	7.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	150.40%#
36) 1,2-Dichloropropane-d6	6.09	67	34723	7.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	151.20%#
41) Toluene-d8	7.33	98	96044	7.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	141.20%#
45) trans-1,3-Dichloropropene-	7.64	79	11323	6.83	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	136.60%#
48) 2-Hexanone-d5	8.11	63	59498	77.20	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	154.40%#
59) 1,1,2,2-Tetrachloroethane-	10.24	84	24849	7.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	153.20%#
65) 1,2-Dichlorobenzene-d4	11.65	152	31950	7.61	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	152.20%#

Target Compounds

					Ovalue
3) Chloromethane	1.25	50	1414	0.256 ug/L	86
12) 1,1-Dichloroethene	2.13	96	11872	3.333 ug/L #	51
25) Chloroform	4.40	83	1892	0.253 ug/L	86
29) 1,1,1-Trichloroethane	4.62	97	4655	0.754 ug/L	97
34) Trichloroethene	5.94	95	1223	0.301 ug/L	94
49) Tetrachloroethene	7.99	164	2113	0.709 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060120\
Data File : VV016411.D
Acq On : 01 Jun 2020 14:54
Operator : SY/MD
Sample : L2780-16
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AL9

Quant Time: Jun 02 03:03:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053020WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 02 01:13:11 2020
Response via : Initial Calibration

